

AD _____

CONTRACT NUMBER: DAMD17-94-C-4081

TITLE: Fluorescence Optic Fiber Stereotactic Needle Ratiometer
for Breast Tumor Diagnosis

PRINCIPLE INVESTIGATOR: Guichen C. Tang, M.D.

CONTRACTING ORGANIZATION: Mediscience Technology Corporation
Cherry Hill, New Jersey 08003

REPORT DATE: October 1996

TYPE OF REPORT: Annual

PREPARED FOR: U.S. Army Medical Research and Materiel Command
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for public release;
distribution unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

19970610 032

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
<small>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing the burden, to Washington Headquarters Service, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.</small>				
1. AGENCY USE ONLY (Leave blank)	2. REPORT DATE October 1996	3. REPORT TYPE AND DATES COVERED Annual (23 Sep 95 - 22 Sep 96)		
4. TITLE AND SUBTITLE Fluorescence Optic Fiber Stereotactic Needle Ratiometer for Breast Tumor Diagnosis		5. FUNDING NUMBERS DAMD17-94-C-4081		
6. AUTHOR(S) Guichen C. Tang, M.D.				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Mediscience Technology Corporation Cherry Hill, New Jersey 08003		8. PERFORMING ORGANIZATION REPORT NUMBER		
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research and Materiel Command Fort Detrick, Maryland 21702-5012		10. SPONSORING / MONITORING AGENCY REPORT NUMBER		
11. SUPPLEMENTARY NOTES				
12a. DISTRIBUTION / AVAILABILITY STATEMENT Approved for Public Release; distribution unlimited		12b. DISTRIBUTION CODE		
13. ABSTRACT (Maximum 200 words) Activities during the second year were still restricted to gaining head start knowledge of the intended instrument from the development of a similar instrument for another project, an aerodigestive cancer diagnosis clinical study at Memorial Sloan-Kettering (MSK). An extension has been granted and a second extension will be requested based on advice from the reviewers at the recent site visit. Patient trials could not be started with just the approval from the Institutional Review Board (IRB) of Mass. Gen. Hosp., but require an Investigational Device Exemption (IDE) from FDA for this Phase I feasibility study. This was determined from the "Significant Risk" ruling by FDA shortly after the award of the Contract which also halted the the Phase II MSK study, both pending an IDE. The Sponsor has engaged the services of an experienced regulatory compliance consultant in drafting our IDE applications. Armed with recent favorable findings from our animal safety studies, the Phase II IDE application is soon to be filed followed by the Phase II application to permit the clinical trials for the device assembled under this contract. Ide approval is expected in 2Q97. In the interim, the device will be assembled, and the range finding data will begin to be collected in vitro on excised tissue. Preliminary baseline data already has been collected on breast tissue (fat, ductile, etc.) using related spectral fluorescence devices in order to develop the algorithm for data analysis. This work will be discussed.				
14. SUBJECT TERMS Breast Cancer Fluorescence, Fiber, Optic, Ratiometer, Needle, Biopsy, Tumor, Cancer		15. NUMBER OF PAGES 11		
		16. PRICE CODE		
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT Unlimited	

FOREWORD

Opinions, interpretations, conclusions and recommendations are those of the author and are not necessarily endorsed by the U.S. Army.

✓ Where copyrighted material is quoted, permission has been obtained to use such material.

✓ Where material from documents designated for limited distribution is quoted, permission has been obtained to use the material.

✓ Citations of commercial organizations and trade names in this report do not constitute an official Department of Army endorsement or approval of the products or services of these organizations.

___ In conducting research using animals, the investigator(s) adhered to the "Guide for the Care and Use of Laboratory Animals," prepared by the Committee on Care and use of Laboratory Animals of the Institute of Laboratory Resources, National Research Council (NIH Publication No. 86-23, Revised 1985).

✓ ___ For the protection of human subjects, the investigator(s) adhered to policies of applicable Federal Law 45 CFR 46.

___ In conducting research utilizing recombinant DNA technology, the investigator(s) adhered to current guidelines promulgated by the National Institutes of Health.

___ In the conduct of research utilizing recombinant DNA, the investigator(s) adhered to the NIH Guidelines for Research Involving Recombinant DNA Molecules.

___ In the conduct of research involving hazardous organisms, the investigator(s) adhered to the CDC-NIH Guide for Biosafety in Microbiological and Biomedical Laboratories.

...



PI - Signature

10/31/86
Date

Table of Contents

Front Cover.....	1
SF298.....	2
Foreword.....	3
Table of Contents.....	4
Introduction.....	5
Body.....	6
Conclusion.....	7
References.....	8
Table I.....	9
Table II.....	10
Figure I.....	11

5. Introduction

The Optic Fiber Stereotactic Needle-based Fluorescence Ratiometer for Breast Tumor Diagnosis (CD-Ratiometer) is an approach to improve the diagnosis of tissue, which is based on fluorescence spectroscopy. The goal of this project is to develop a CD-Ratiometer equipped with a small diameter optic fiber suitable for insertion into a hollow metallic needle for *in vivo* breast tumor diagnosis. A compact bread board CD-Ratiometer can measure the ratio of the fluorescence intensities at two pre-determined wavelengths, emitted from human tissue which has been photoexcited by a third pre-determined wavelength. This unit will be tested at Massachusetts General Hospital. The CD-Ratiometer can give information in real time concerning the condition of the tissue, information that normally is obtained from biopsy and pathology, thereby, possibly permitting the physician to make medical judgments more quickly. The device, through its needle-fiber optic probe, delivers an excitation light beam to a breast tumor, collects fluorescence emitted from the tumor and delivers this light back through a set of narrow band filters to a pair of photo-detectors for data analysis.

Researchers at the Mediphotonics Laboratory (MPL) at the City College of New York (CCNY) supported by Mediscience Technology Corp. (MTC) have been continuing to develop spectroscopic techniques to evaluate, diagnose and characterize tissues. Differences between malignant and normal or benign *in vitro* breast [1], gynecological [2] and colon [3] tissue samples were observed and ratios of their spectral intensities at various key wavelengths were determined.

During the report period, no work was performed on the Ratiometer under this contract because execution of the subcontract with City College of New York (CCNY) of CUNY was delayed until September 1st pending completion of the revisions of patient Consent Form required by the Human Use review and regulatory Affairs Division, USAMRMC, and resolution of indemnification issues raised by Research Foundation of CUNY. Contract extensions were obtained.

Recent research on breast tissues [4] demonstrated that the diagnostic accuracy would be improved using multiple wavelengths for tissue measurements as criteria to separate the normal and abnormal tissues. Therefore, an updated algorithm from these results [4] will be incorporated for tissue diagnostics using the Ratiometer unit (Fig.1).

6. Body

Regulatory Compliance Matters

Planned testing on human subjects for this contract had to be postponed due to a ruling by the FDA in January 1995 involving a similar Mediscience device for another clinical study but applicable also to this study. That target organ was the oral cavity. That ruling required the Company to conduct additional toxicology studies in an animal model before the FDA would consider granting an Investigation Device Exemption ("IDE") to conduct in-vivo clinical studies on humans. The ruling designated that fluorescence spectral device and, by inference, all such devices as "Significant Risk" as opposed to "Nonsignificant Risk".

The additional regulatory compliance burden placed on the Company by the FDA has necessitated the purchase of specialized regulatory affairs services by Mediscience for developing, drafting, and negotiating/defending the appropriateness of the required IDE application(s) with the FDA. During the second year of this contract, the toxicology study described above was completed. Data collected initially from human oral cavity study prior to the FDA ruling is currently in the process of its assemblage into the clinical study report as a part of a Phase II IDE resubmission on or about November 1, 1996.

Mediscience expects the "Breast Needle Ratiometer" to benefit from the favorable results of the toxicology study. The company expects to file for the "Breast Needle Ratiometer" Phase I IDE with the FDA by April 1, 1997, and prior to that time it expects to have completed the assembly of the "Breast Needle Ratiometer". This work has already begun. Also, final range finding studies with the needle probe will have begun from ex-vivo optical ratio analysis of freshly excised breast tumor tissue. Preliminary ex-vivo range finding studies, not using the needle probe, were already successfully concluded. Hence, the in-vivo human clinical feasibility study can begin immediately after the FDA approves the "Breast Needle Ratiometer" IDE.

Other regulatory driven activities begun that are related to this contract include the purchase of design engineering services to properly document the "Breast Needle Ratiometer" device with drawings, parts and assembly specifications and procedures to fulfill FDA Good Manufacturing Practice requirements. Additionally, the Company will now provide an on-site field service engineer to support its clinical studies and is further required by the FDA to provide for professional study monitoring and data management by an independent clinical research organization.

Summary of Activities

Because the subcontract with CCNY was delayed, the PI was reassigned to other projects. However, related research, not under the subcontract, continued at CCNY. A cancer diagnosis procedure [4] was developed as summarized by the flow chart, Fig.1. This is reflected in the design of the filter wheel and planned fabrication of the unit. Also, during this period, optic fiber and needle acquisitions were made and patient consent and subcontract issues were resolved

The work plan for the upcoming year is :

- (1) Subcontract to CCNY to assemble a needle ratiometer unit
- (2) Measurements on *in vitro* tissues including breast sample at CCNY and MGH.
- (3) Testing *in vivo* tissues at MGH pending IDE approval (MTC)
- (4) Final report

The plan for the next period is shown in the table II.

7. Conclusion

The project's parameters have expanded from the original plan owing to the IDE requirements. This motivated our request for the twelve month extension that was granted. However, that extension will carry the project into just the start of the human clinical feasibility trials. Therefore, a second twelve month extension was requested and granted to permit the completion and assessment of the clinical feasibility trials. Positive findings would provide the basis for expanded clinical trials and a Phase II IDE application to FDA. During this period, the algorithm for wavelength pairs will be refined for use in the needle ratiometer unit.

References

1. R.R.Alfano, B.B.Das, J.Cleary, R.Prudente, E.J.Celmer, Light sheds light on cancer-distinguishing malignant tumor from benign tissue and tumors, Bull. N.Y. Acad. Med., second series 67(2), 143 (1991).
2. W.S.Glassman, C.H.Liu, G.C.Tang, S.Lubicz, R.R.Alfano, Ultraviolet excited fluorescence spectra from non-malignant and malignant tissues of gynecological tract, Lasers in the Life Science 5(1-2), 49-58 (1992).
3. Y.L. Yang, G.C. Tang, M. Bessler, and R.R. Alfano, Fluorescence spectroscopy as a photonic pathology method for detecting colon cancer, Lasers in the Life Science, Vol. 6(4), 259-276 (1995).
4. Y. L. Yang, A. Katz, E.J. Edward, J. Celmer, M. Z. Szczepaniak, and R.R. Alfano, Optical spectroscopy of benign and malignant breast tissues, Lasers in the Life Science, Vol. 7(2), 115-127, (1996).

Table I, Filter selection for Fiber Needle-based Fluorescence Ratiometer

For Fluorescence mode		
Excitation Wavelength	Fluorescence Wavelength	Fluorescence Wavelength
300 nm	340 nm	440 nm
For Synchronous mode		
Excitation Wavelength Pair		Probe Wavelength Pair
520/460 (nm)		520/460 (nm)
Excitation Wavelength		Probe Wavelength
420/460 (nm)		420/460 (nm)

Table II Project Schedule

Tasks	Time Table (Months)												
	0	1	2	3	4	5	6	7	8	9	10	11	12
	(Oct.	Nov.	Dec.	Jan.	Feb.	Mar.	Apr.	May	Jun	July	Aug.	Sept.	
Order parts	0												
Optics													
Electronics													
Mechanics													
Software enhancement	0												
Unit assembly				0									
Enhance needle design	0												
Test unit operation				0									
Test in vitro tissues				0									
Hospital test in vitro						0							
IDE approval for in vivo test (?)									0				
Report									0				

**Fig. 1 Flow Chart of Cancer detection
Using Synchronous Scan and fluorescence**

